

From Addendum 5 is the following:

Resident Room Plans for rooms 1119, 1219, 2119, 2219, 2319 and 2311 to make these rooms **Bariatric Rooms** complete with ceiling mounted patient lift system. Basis for Design of the lift system to be **the ARJO Maxi Sky 2 600# lifting capacity system**. System to be a complete installation with all tracks, intermediate supports at the building structure for the track, controllers, slings, and other customary appurtenances for a typical installation. Each room to have a lift system, not portable.